

52252-7

9/25/2013

1/13



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
WASHINGTON, D.C. 20460

SEP 25 2013

OFFICE OF CHEMICAL SAFETY
AND POLLUTION PREVENTION

Richard Ormsbee, Corporate Regulatory Affairs Manager
Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447-4822

Subject: Notification; Per PR-Notice 98-10
Product Name: Actril Cold Sterilant
EPA Registration Number: 52252-7
Application Receipt Date: August 27, 2013

Dear Mr. Ormsbee:

The Agency is in receipt of your Application for Pesticide Notification under Pesticide Registration Notice (PRN) 98-10 dated August 26, 2013 for the product referenced above. The Antimicrobials Division (AD) has conducted a review of this request for its applicability under PRN 98-10 and finds that the action requested falls within the scope of PRN 98-10. The updated label with the addition of the revised company name, contact information, and fogging language will be placed in our records.

If you have any questions regarding this letter, please contact Amaris Johnson at (703) 305-9542 or johnson.amaris@epa.gov.

Sincerely,

A handwritten signature in black ink that reads "M Swindell".

Marshall Swindell
Product Manager 33
Regulatory Management Branch 1
Antimicrobial Division (7510P)
Office of Pesticide Programs

2/13

Please read instructions on reverse before completing form.

Form Approved, OMB No. 2070-0060

Print Form



United States
Environmental Protection Agency
Washington, DC 20460

<input type="checkbox"/>	Registration
<input type="checkbox"/>	Amendment
<input checked="" type="checkbox"/>	Other

OPP Identifier Number

Application for Pesticide - Section I

1. Company/Product Number 52252-7	2. EPA Product Manager Marshall Swindell	3. Proposed Classification <input checked="" type="checkbox"/> None <input type="checkbox"/> Restricted
4. Company/Product (Name) Acril Cold Sterilant	PM# 33	
5. Name and Address of Applicant (Include ZIP Code) Medivators Inc 14605 28th Avenue North Minneapolis, MN 55447 <input type="checkbox"/> Check if this is a new address		6. Expedited Review. In accordance with FIFRA Section 3(c)(3) (b)(i), my product is similar or identical in composition and labeling to: <input checked="" type="checkbox"/> EPA Reg. No. _____ Product Name _____

Section - II

<input type="checkbox"/> Amendment - Explain below.	<input type="checkbox"/> Final printed labels in response to Agency letter dated _____
<input type="checkbox"/> Resubmission in response to Agency letter dated _____	<input type="checkbox"/> "Me Too" Application.
<input checked="" type="checkbox"/> Notification - Explain below.	<input type="checkbox"/> Other - Explain below.

Explanation: Use additional page(s) if necessary. (For section I and Section II.)

See attached additional page.

Section - III

1. Material This Product Will Be Packaged In:				2. Type of Container	
Child-Resistant Packaging <input type="checkbox"/> Yes* <input type="checkbox"/> No	Unit Packaging <input type="checkbox"/> Yes <input type="checkbox"/> No	Water Soluble Packaging <input type="checkbox"/> Yes <input type="checkbox"/> No	Metal Plastic Glass Paper Other (Specify) _____		
* Certification must be submitted		If "Yes" Unit Packaging wgt. No. per container	If "Yes" Package wgt. No. per container		
3. Location of Net Contents Information <input type="checkbox"/> Label <input type="checkbox"/> Container		4. Size(s) Retail Container		5. Location of Label Directions <input type="checkbox"/> On Label <input type="checkbox"/> On Labeling accompanying product	
6. Manner in Which Label is Affixed to Product <input type="checkbox"/> Lithograph <input type="checkbox"/> Paper glued <input type="checkbox"/> Stenciled			<input type="checkbox"/> Other _____		

Section - IV

1. Contact Point (Complete items directly below for identification of individual to be contacted, if necessary, to process this application.)		
Name Richard M. Ormsbee	Title Manager, RA	Telephone No. (Include Area Code) 763-551-2689
Certification I certify that the statements I have made on this form and all attachments thereto are true, accurate and complete. I acknowledge that any knowingly false or misleading statement may be punishable by fine or imprisonment or both under applicable law.		6. Date Application Received (Stamped)
2. Signature 	3. Title Manager, RA	
4. Typed Name Richard M. Ormsbee	5. Date 8/26/2013	

9/13

**ACTRIL® COLD STERILANT
Master Label**

For Use In Health Care Environments

Active Ingredients

Hydrogen Peroxide 1.00%
Peroxyacetic Acid08%

Inert Ingredients 98.92%

Total 100.00%

#52252-7
NOTIFICATION
Date Reviewed: 9/25/13
Reviewed By: *A. Johnson*

EPA Reg. No. 52252-7

EPA Est. No. 52252-MN-01

Net Contents: _____

KEEP OUT OF REACH OF CHILDREN

DANGER - PELIGRO

See {side panel} {Directions for Use} for additional precautionary statements

Lot No. _____

Expiration Date: _____

(Medivators logo)

Medivators Inc. Headquarters
14605 28th Avenue North
Minneapolis, MN 55447 U.S.A.
Toll Free: (800) 444-4729
www.Medivators.com

5/13

PRECAUTIONARY STATEMENTS

Hazard to Humans and Domestic Animals.

DANGER - PELIGRO

Corrosive. Causes irreversible eye damage. Harmful if absorbed through skin. Do not get in eyes, on skin, or on clothing. Avoid contact with skin. Wash hands before eating, drinking, chewing gum, using tobacco or using the toilet. Remove contaminated clothing and wash hands before reuse. Caution should be used when applying indoors because pets may be at risk.

FIRST AID	
IF INHALED	<ul style="list-style-type: none"> • Move person to fresh air. • If person is not breathing, call 911 or an ambulance, then give artificial respiration, preferably mouth-to-mouth, if possible. • Call a poison control center or doctor for further treatment advice.
IF ON SKIN OR ON CLOTHING	<ul style="list-style-type: none"> • Take off contaminated clothing. • Rinse skin immediately with plenty of water for 15-20 minutes. • Call a poison control center or doctor for treatment.
IF IN EYES	<ul style="list-style-type: none"> • Hold eye open and rinse slowly and gently with water for 15-20 minutes. • Remove contact lenses, if present, after first 5 minutes, then continue rinsing. • Call a poison control center or doctor for treatment advice.
IF SWALLOWED	<ul style="list-style-type: none"> • Call a poison control center or doctor immediately for treatment advice. • Have person sip a glass of water if able to swallow. • Do not induce vomiting unless told to by a poison control center or doctor. • Do not give anything by mouth to an unconscious person.
HOT LINE NUMBER	
For chemical emergency, spill, leak, fire, exposure and accident, call Chemtrec, day or night (800) 424-9300, (703) 527-3387.	
Have the product container or label with you when calling a poison control center or doctor, or going for treatment. You may also contact Chemtrec at 1-800-424-9300 for emergency medical treatment information.	
Note to Physician: Probable mucosal damage may contraindicate the use of gastric lavage.	

PRECAUTIONARY STATEMENTS Cont.

Personal Protective Equipment (PPE)

Handlers must wear: Goggles or face shield, and protective rubber gloves.

Environmental Hazards

This pesticide is toxic to birds, fish and aquatic invertebrates.

(Applicable to volumes of 5 gallons or greater.)

Do not discharge effluent containing this product into lakes, streams, ponds, estuaries, oceans, or public waters unless this product is specifically identified and addressed in a NPDES permit. Do not discharge effluent containing this product to sewer systems without previously notifying the sewage plant authority. For guidance contact your State Water Board or Regional Office of the U.S. Environmental Protection Agency.

Physical and Chemical Hazards

This product contains an oxidizing agent.

PRODUCT CLAIMS

- Sterilant undiluted 5½ hr. 20°C
- Sporicidal undiluted 5½ hr. 20°C
- Bactericidal undiluted 5½ hr. 20°C
- Virucidal undiluted 5½ hr. 20°C
 - Hepatitis B undiluted 5½ hr. 20°C
 - HIV-1* undiluted 5½ hr. 20°C
- Fungicidal undiluted 5½ hr. 20°C
- Broad Spectrum Disinfectant undiluted 10 min. 20°C
- Hospital Disinfectant Dilute 50X 10min. 20°C
 - and Pseudomonacidal Use Fresh Solution Only, Do Not Reuse
- Cleaner-Sanitizer 50X dilution 5 min Do Not Reuse
 - (non-food contact surfaces)
- Tuberculocidal undiluted 10 min. 20°C
 - Use Fresh Solution Only, Do Not Reuse
- Non-Staining
- Pre-Activated/No Activation Needed
- Fast-Acting
- Reusable
 - As a sterilant for 14 days
 - As a broad spectrum disinfectant for 14 days (when used undiluted)
- Germicidal Spray Disinfectant undiluted 30 sec. 20°C

Failure to follow directions for use per package insert may result in user injury.

* KILLS HIV-1 ON PRE-CLEANED ENVIRONMENTAL SURFACES/OBJECTS PREVIOUSLY SOILED WITH BLOOD/BODY FLUIDS in health care settings or other settings in which there is an expected likelihood of soiling inanimate surfaces/objects with blood or body fluids, and in which the surfaces/objects likely to be soiled with blood or body fluids can be associated with the potential for transmission of HIV-1 (Human Immunodeficiency Virus Type 1 - associated with AIDS).

SPECIAL INSTRUCTIONS FOR CLEANING AND DECONTAMINATION AGAINST HIV OF SURFACES/OBJECTS SOILED WITH BLOOD/BODY FLUIDS:

PERSONAL PROTECTION: Specific barrier protection items to be used when handling items soiled with blood or body fluids are disposable latex gloves, gowns, masks and eye coverings.

CLEANING PROCEDURE: Blood and other body fluids must be thoroughly cleaned from surfaces and objects before the application of the sterilant.

DISPOSAL OF INFECTIOUS MATERIALS: Blood and other body fluids should be autoclaved and disposed of according to federal, state, and local regulations for infectious waste disposal.

CONTACT TIME: Leave surfaces wet with undiluted Actril Cold Sterilant for a minimum of 5½ hours

DIRECTIONS FOR USE

1. It is a violation of Federal Law to use this product in a manner inconsistent with its labeling.
2. Do not use after Expiration Date.
3. Do not allow Actril Cold Sterilant to mix with alkaline substances such as bleach (Sodium hypochlorite) or other oxidizing agents.
4. Use purified water (e.g. deionized) for making dilutions, cleaning and rinsing.
5. Some materials may be incompatible with Actril Cold Sterilant. Test material prior to use.
6. Reuse of diluted Actril Cold Sterilant is not recommended.

DIRECTIONS FOR STERILIZATION

1. Actril Cold Sterilant may be used to sterilize non-invasive dental/surgical equipment such as pressure clamps, tubing clamps, blood pressure cuffs, mirrors etc. This product is not to be used as a terminal sterilant on any critical/semi-critical medical device.
2. Remove any obvious debris or organic material from the surface to be sterilized. This can often be accomplished by rinsing with water or by detergent cleaning followed by a water rinse.
3. Immerse the item to be sterilized in a sufficient volume of undiluted Actril Cold Sterilant to cover the item and fill all passages requiring sterilization. Hold in sterilizing solution for a minimum of 5½ hours at 20°C temperature (68°F).
4. Remove items after 5½ hours and rinse with sterile water until rinse water shows level of 10 ppm or less when tested with Actril Residual Test Strips.
5. The solution may be used and reused for up to 14 days in a manual system with 5½ hours immersion.

DIRECTIONS FOR BROAD SPECTRUM DISINFECTION

1. For broad spectrum disinfection of items such as external surfaces of medical equipment, kidney dialysis machines, non-invasive dental instruments, counters, bathroom fixtures, etc. This product is not to be used as a terminal high level disinfectant on any critical/semi-critical medical device.
2. Remove any obvious debris or organic material from the surface to be disinfected. This can often be accomplished by rinsing with purified water (e.g. deionized), mechanical action, or by detergent cleaning followed by a water rinse.
3. Immerse the item to be disinfected in sufficient volume of undiluted Actril Cold Sterilant solution to cover the item and/or where appropriate by filling all passages requiring disinfection. Hold item in contact with the disinfecting solution for a minimum of 10 minutes at 20°C temperature (68°F).
4. Remove items after 10 minutes and rinse with purified water until effluent shows acceptable levels when tested with Actril Residual Test Strips.
5. The solution may be used and reused for up to 14 days in a manual system with 10 minutes immersion.
6. For tuberculocidal activity at 20°C with fresh, undiluted solution, immerse completely for 10 minutes. Remove items and rinse thoroughly.
7. This product may be used to pre-clean or decontaminate critical or semi-critical medical devices prior to sterilization or high level disinfection.

9/13

DIRECTIONS FOR HOSPITAL DISINFECTION

1. For disinfection of items such as external surfaces of medical equipment, counters, bathroom fixtures, etc. This product is not to be used as a terminal high level disinfectant on any critical/semi-critical medical device.
2. Remove any obvious debris or organic material from the surface to be disinfected. This can often be accomplished by rinsing with purified water or by detergent cleaning followed by a water rinse.
3. Immerse the pre-cleaned item to be disinfected in sufficient volume of freshly diluted 50X Actril Cold Sterilant using purified water, to cover the item and fill all passages requiring disinfection. Hold item in contact with the disinfecting solution for a minimum of 10 minutes at 20°C temperature (68°F).
4. Remove items after 10 minutes and rinse with purified water until rinse water shows levels of 10 ppm or less when tested with Actril Residual Test Strips.
5. Cannot be reused as a hospital disinfectant.
6. This product may be used to pre-clean or decontaminate critical or semi-critical medical devices prior to sterilization or high level disinfection.

FOGGING FOR REGULAR CLEANING:

This product may be used for fogging (wet misting) following regular cleaning procedure for hard room surfaces.

1. Prior to fogging, remove or carefully protect all food products and packaging materials.
2. Ensure room is properly ventilated. Vacate all personnel from the room during fogging and for a minimum of 2 hours after fogging or until the hydrogen peroxide air concentration is below 0.5 ppm.
3. Fog areas using one quart (946 ml) per 1000 cu. ft. (28.3 m³) of room volume with undiluted Actril solution.
4. Allow surfaces to dry thoroughly before operations are resumed.

Note: In all applications, always prepare a new solution daily to ensure effectiveness. Do not reuse solutions. Dispose of any unused solution.

Directions for Use in the AxSYM/AxSYM Plus Instrument

It is a violation of Federal Law to use this product in a manner inconsistent with its labeling.

This product is to be used only for cleaning and inhibition of fouling bacteria in the AxSYM/AxSYM Plus tubing system of their in-vitro immunoassay analyzer.

For (*in vitro*) diagnostic use only.

For use in the AxSYM SYSTEM for Monthly Maintenance of Tubing Decontamination.

Perform the procedure found in the AxSYM SYSTEM Operations Manual, Section 9: Service and Maintenance. In accordance with the Operations Manual perform the following:

1. Prepare a 5% solution by diluting 20 ml into 380 ml of approved water. Note: The dilution solution is to be used immediately.
2. Empty the Liquid Waste Container and update Liquid Waste Inventory.
3. Select MAINTENANCE from the Main Menu screen.
4. Select TUBING DECONTAMINATION.
5. Select OK and allow the system to run for the entire cycle.
6. Rinse and flush with water per directions on the Tubing Decontamination Rinse screen.

11/13

Directions for Use as a Germicidal Disinfectant Spray

1. Spray Actril Cold Sterilant undiluted onto cleaned surface using a plastic spray bottle.
2. Allow to remain on surface for 30 seconds.
3. Let air dry or rinse with purified water, drain off excess water if possible and allow to dry.

**Directions for Use as a Cleaner/Sanitizer
(Non-food contact surfaces)**

1. Using water or mechanical action, remove heavy soil or gross filth from hard surfaces such as Formica, stainless steel or vinyl surfaces.
2. Apply by cloth, mop or sponge so as to wet all surfaces thoroughly, a freshly made 50X dilution of Actril Cold Sterilant, made using purified water, to the pre-cleaned surface or immerse pre-cleaned items to be sanitized in the solution. Allow 5 minutes of contact time. Let air dry or rinse with purified water, drain excess if possible and allow to air dry.
3. May NOT be reused as a cleaner/sanitizer.

USES

Primarily intended for sterilization, disinfection or sanitization of:

- Stainless Steel
- Plastic Items and surfaces
- Anesthesia Equipment
- Medical Equipment
- External Surfaces of Respiratory Equipment
- Hard Surfaces e.g. counter tops
- Non-Invasive Surgical Equipment
- Floors
- Walls
- Bathroom fixtures
- Glass
- Non-Invasive Dental Instruments
- Formica
- Vinyl

Materials Compatibility: Polyvinyl chloride, polypropylene, polyurethane, aceto-copolymers, polycarbonate, polysulfone, polystyrene, stainless steel.

12/13

STORAGE AND DISPOSAL

Do not contaminate water, food or feed by storage or disposal.

Pesticide Storage

1. Store upright in shipping carton.
2. Do not expose to direct sunlight.
3. Maintain temperature below 75°F (24°C)
4. Avoid contact with combustible materials.
5. Store in original closed container.
6. For chemical emergency, spill, leak, fire, exposure and accident, call Chemtrec, day or night (800) 424-9300, (703) 527-3387.

Pesticide Disposal

Wastes resulting from the use of this product may be disposed of on-site in a sanitary sewer or at an approved waste disposal facility.

Pesticide wastes are acutely hazardous. Improper disposal of excess pesticide, spray mixture, or rinsate is a violation of Federal Law. If these wastes cannot be disposed of by use according to label instructions, contact your State Pesticide or Environmental Control Agency, or the Hazardous Waste Representative at the nearest EPA Regional Office for Guidance.

Container Disposal

Nonrefillable container. Do not reuse or refill this container. Offer for recycling, if available.

(For ≤ 5 gal.): Clean container promptly after emptying. Triple rinse as follows: Empty the remaining contents into application equipment or a mix tank and drain for 10 seconds after the flow begins to drip. Fill container ¼ full with water and recap. Shake for 10 seconds. Pour rinsate into application equipment or a mix tank or store rinsate for later use or disposal. Drain for 10 seconds after flow begins to drip. Repeat this procedure two more times.

(For >5 gal.): Clean container promptly after emptying. Triple rinse as follows: Empty the remaining contents into application equipment or a mix tank. Fill the container ¼ full with water. Replace and tighten enclosures. Tip container on its side and roll it back and forth, ensuring at least one complete revolution, for 30 seconds. Stand container on its end and tip it back and forth several times. Empty rinsate into application equipment or a mix tank or store rinsate for later use or disposal. Repeat this procedure two more times.

{Optional Statements}

Packaged in Clean Room Conditions

6 x 1 liter

6 x 0.96 liter

Packaged with Sprayheads

Double Bagged

ETO Exposed Packaging

ETO Processed Packaging

Ethylene Oxide Exposed Packaging

Ethylene Oxide Processed Packaging